


BeMe Measurement Guide



Child name: _____

Date: _____

Location: _____

Therapist: _____

Notes: _____

- A. Thigh length [Back of knee to back of bottom] _____
- B. Hip Width [From either side of widest part of hips] _____
- C. Back height [From bottom to top of shoulder] _____
- D. Chest width [from either side of chest, nipple level] _____
- E. Lower leg length [From back of knee to underside of foot] _____

Child Size			BeMe Sizing			
			1	2	3	4
Seat Base	Seat Depth [mm]	A	120-270*	225-350	320-430	400-540
	Standard Hip Guide Width [mm]	B	145-215	175-260	245-350	290-405
	Slim Hip Guide Width [mm]	B	170-240	200-285	270-375	315-430
	Seat Width Maximum [No Hip Guides] [mm]	B	280	325	410	475
Back support	Back Height [mm]	C	300-400	350-470	400-530	475-650
	Lateral Width [mm]	D	130-230	200-300	225-350	350-450
	Inset Lateral Width [mm]	D	80-180	150-250	175-300	300-400
Lower limb support	Lower Leg Length [mm]	E	120-270	225-350	320-440	400-540
Weight limit	Max Weight Limit **		35	50	70	80

*120 to 170 option requires spacer, 170 or below not compatible with contoured seat options

** System max user weight should be lesser of seat, chassis and backrest